



**For the use of Sarstedt consumables
in the field of cell and tissue culture**

TC Tested items fulfil the following certified purity criteria:

- ✓ Sterile
- ✓ Pyrogen-free/endotoxin-free
- ✓ Non-cytotoxic
- ✓ DNA-free
- ✓ DNase/RNase-free

We guarantee that the following limit values are met:

Sterility validated according to ISO 11137

Pyrogens/endotoxins <0.06 EU/ml

Non-cytotoxic according to ISO 10993

Human DNA <0.5 pg/μl

Bacterial DNA <0.02 pg/μl

DNase <7.1x10⁻⁵ U/μl

RNase <7.1x10⁻⁹ U/μl