



For the use of Sarstedt consumables in nucleic acid analytics

PCR Performance Tested items fulfil the following purity criteria certified by an independent laboratory:

- ✓ DNA-free
- ✓ DNase/RNase-free
- ✓ PCR inhibitor-free

We guarantee that the following limit values are met:

Human DNA <0.5 pg/μl

Bacterial DNA <0.02 pg/μl

DNase <1x10⁻⁵ U/μl

RNase <1x10⁻⁹ Kunitz units/μl

Real Time PCR <0.5 cycles Ct value shifting compared to control